


| | | | | | |
|---|--|---|--------------------|------------------|--|
| Patient ID SA00118475 | Patient Name VALIDATIONTEST, FIBRO | Birth Date 1970-02-27 | Gender M | Age 49 | |
| Order Number SA00118475 | Client Order Number SA00118475 | Ordering Physician CLIENT, CLIENT | Report Notes | | |
| Account Information C7028846 DLMP Rochester | | Collected 28 Feb 2019 00:00 | | | |


FibroTest Score MCR | **FibroTest Stage** MCR
 0.03 | F0



FibroTest Interpretation MCR
 no fibrosis
 FibroTest estimates liver fibrosis

| FibroTest Score | Stage | Interpretation |
|-----------------|-------|-----------------------------|
| 0.00–0.21 | F0 | no fibrosis |
| 0.21–0.27 | F0-F1 | no fibrosis |
| 0.27–0.31 | F1 | minimal fibrosis |
| 0.31–0.48 | F1-F2 | minimal fibrosis |
| 0.48–0.58 | F2 | moderate fibrosis |
| 0.58–0.72 | F3 | advanced fibrosis |
| 0.72–0.74 | F3-F4 | advanced fibrosis |
| 0.74–1.00 | F4 | severe fibrosis (Cirrhosis) |

ActiTest Score MCR | **ActiTest Grade** MCR
 0.05 | A0



ActiTest Interpretation MCR
 no activity
 ActiTest estimates necroinflammatory activity

| ActiTest Score | Grade | Interpretation |
|----------------|-------|----------------------|
| 0.00–0.17 | A0 | no activity |
| 0.17–0.29 | A0-A1 | no activity |
| 0.29–0.36 | A1 | minimal activity |
| 0.36–0.52 | A1-A2 | minimal activity |
| 0.52–0.60 | A2 | significant activity |
| 0.60–0.62 | A2-A3 | significant activity |
| 0.62–1.00 | A3 | severe activity |

FibroTest-ActiTest Comment MCR

The reliability of results is dependent on compliance with the preanalytical and analytical conditions recommended by BioPredictive. The tests have to be deferred for: acute hemolysis, acute hepatitis, acute inflammation, extra hepatic cholestasis. The advice of a specialist should be sought for interpretation in chronic hemolysis and Gilbert's syndrome. The test interpretation is not validated in liver transplant patients. Isolated extreme values of one of the components should lead to caution in interpreting the results. In case of discordance between a biopsy result and a test, it is recommended to seek advice of a specialist. The causes of these discordances could be due to a flaw of the test or to a flaw in the biopsy: i.e. a liver biopsy has a 33% variability rate for one fibrosis stage. FibroTest is interpretable for chronic hepatitis B and C, alcoholic and non alcoholic steatosis. ActiTest is interpretable for chronic hepatitis B and C.

BioPredictive Serial Number MCR

148534

Performing Site Legend

| Code | Laboratory | Address | Lab Director | CLIA Certificate |
|------|---|--|-----------------------------|------------------|
| MCR | Mayo Clinic Laboratories - Rochester Main Campus | 200 First Street SW, Rochester, MN 55905 | William G. Morice M.D. Ph.D | 24D0404292 |
| SDL | Mayo Clinic Laboratories - Rochester Superior Drive | 3050 Superior Drive NW, Rochester MN 55901 | William G. Morice M.D. Ph.D | 24D1040592 |

| | | | | |
|---|--|---|--------------------|------------------|
| Patient ID SA00118475 | Patient Name VALIDATIONTEST, FIBRO | Birth Date 1970-02-27 | Gender M | Age 49 |
| Order Number SA00118475 | Client Order Number SA00118475 | Ordering Physician CLIENT, CLIENT | Report Notes | |
| Account Information C7028846 DLMP Rochester | | Collected 28 Feb 2019 00:00 | | |

FibroTest-ActiTest, S

Apolipoprotein A1, S
181 mg/dL
MCR
Reference Value
≥120
Alanine Aminotransferase (ALT), S
18 U/L
MCR
Reference Value
7–55
Alpha-2-Macroglobulin, S
141 mg/dL
SDL
Reference Value
100–280
Gamma Glutamyltransferase (GGT), S
11 U/L
MCR
Reference Value
8–61
Haptoglobin, S
166 mg/dL
SDL
Reference Value
30–200
Bilirubin, Total, S
0.3 mg/dL
MCR
Reference Value
≤1.2
Received: 28 Feb 2019 14:33

Reported: 28 Feb 2019 14:37

Laboratory Notes

- 1 This test was developed and its performance characteristics determined by Mayo Clinic in a manner consistent with CLIA requirements. This test has not been cleared or approved by the U.S. Food and Drug Administration.

Performing Site Legend

| Code | Laboratory | Address | Lab Director | CLIA Certificate |
|------|---|--|-----------------------------|------------------|
| MCR | Mayo Clinic Laboratories - Rochester Main Campus | 200 First Street SW, Rochester, MN 55905 | William G. Morice M.D. Ph.D | 24D0404292 |
| SDL | Mayo Clinic Laboratories - Rochester Superior Drive | 3050 Superior Drive NW, Rochester MN 55901 | William G. Morice M.D. Ph.D | 24D1040592 |